

Fig. 1

Calculation of Laboratory Mean

Run ID	Result	Cum. Mean	% diff (n=30)	SDI	Run ID	Result	Cum. Mean	% diff (n=30)	SDI
1	12.4	12.4	-4.6	-0.6	16	11.9	11.9	-0.2	0.0
2	11.0	11.7	1.1	0.1	17	11.1	11.8	0.2	0.0
3	11.2	11.5	2.6	0.4	18	11.3	11.8	0.5	0.1
4	11.6	11.5	2.6	0.3	19	12.5	11.8	0.1	0.0
5	14.2	12.1	-1.9	-0.3	20	10.4	11.8	0.7	0.1
6	11.8	12.0	-1.5	-0.2	21	12.1	11.8	0.6	0.1
7	11.6	12.0	-1.1	-0.1	22	12.0	11.8	0.5	0.1
8	11.9	12.0	-1.0	-0.1	23	11.2	11.8	0.7	0.1
9	12.5	12.0	-1.5	-0.2	24	11.8	11.8	0.7	0.1
10	11.1	11.9	-0.8	-0.1	25	11.6	11.8	0.8	0.1
11	10.4	11.8	0.4	0.1	26	11.3	11.7	0.9	0.1
12	12.1	11.8	0.2	0.0	27	13.9	11.8	0.3	0.0
13	12.4	11.9	-0.2	0.0	28	11.6	11.8	0.3	0.0
14	12.1	11.9	-0.3	0.0	29	13.6	11.9	-0.2	0.0
15	11.6	11.9	-0.1	0.0	30	11.2	11.8	0.0	0.0
								Mean	11.8
								SD	0.9
								CV	7.9
								Min	10.4
								Max	14.2

FIGURES/COURTESY DAVID PLAUT