

## Instrument Validation and Implementation Record

Instrument:

Model

Serial Number

Item	Date	Person	Signature
Date Installed			
Location:			
Initial Validation by Vendor			
Staff Training by Vendor			
Procedure Manual in Place			
Supervisor Approval			
Medical Director Approval			
Maintenance Record in Place			
Supervisor Approval for Instrument Use			
Medical Director Approval for instrument Use			

Notes:

## Instrument Validation and Implementation Record

Instrument: Leica Peloris Tissue Processor  
 Model: Peloris 1  
 Serial Number: 0265185B/PEL10

Item	Date	Person	Signature
Date Installed	12/17/2012	Leica	
Location: Histology lab room B234, Mt Zion	N/A	N/A	N/A
Initial Validation by Vendor <i>please see note below</i>	12/17/2012	Leica	<i>E 2</i>
Staff Training by Vendor	12/17/2012	Leica	<i>E 2</i>
Procedure Manual in Place: On instrument	12/17/2012	Leica	<i>E 2</i>
Supervisor Approval	12/17/2012	Maria Kallala	<i>Maria</i>
Medical Director Approval	12/17/2012	Andrew Horvath	<i>Andrew</i>
Maintenance Record in Place	12/17/2012	Maria Kallala	<i>Maria</i>
Supervisor Approval for Instrument Use	12/17/2012	Maria Kallala	<i>Maria</i>
Medical Director Approval for instrument Use	12/17/2012	Andrew Horvath	<i>Andrew</i>

**Notes:** Instrument was installed as a demo on Dec 17, 2012 then purchased on 12/26/2012

*Note regarding validation - upon installation instrument performed to manufacturer specifications*

## Special Stain Validation

Date: 8-17-2011

Special Stain: Warthin-Starry

Check one: ☐ Stain Validation

☒ Control Slide Validation

Is this special stain comparable to the current manual method and suitable for general use?  
(Circle one) ☒ YES    NO

Approved by: \_\_\_\_\_

If NO, explain inadequacy:

Date	Test Case	Tissue Type	Manual Result (check/circle one)	Comments	Suitable to use for staining?
8-17-2011	Current control slide	W-S CSS01	Inadequate <input checked="" type="checkbox"/> Satisfactory	Positive, but scant organisms	Yes
8-17-2011	"Spirochete-2"	Liver	Inadequate <input checked="" type="checkbox"/> Satisfactory	Numerous thin spirochetes	Yes
8-17-2011	"H. pylori"	Stomach	Inadequate <input checked="" type="checkbox"/> Satisfactory	Numerous organisms	Yes
8-17-2011	ZS07-8828	Colon	Inadequate <input checked="" type="checkbox"/> Satisfactory	Positive lawn of spirochetes on mucosa	Control only
8-17-2011	W-S Negative control		Inadequate Satisfactory N/A	Negative	Negative control
			Inadequate Satisfactory		

**Equipment Validation  
VIP5 Unit "C"**

**Model Number: VIP5A-B1**  
**Serial Number: 52140113**  
**Installed 6/28/2008**

**Clinical Engineering Number: 040252**  
**Electrical Check Done (115 VAC  $\pm$  10%, 60 HZ, 11.0A): 7/2007**  
**Operating Conditions: Ambient Temperature: 10 C – 40 C**  
**Relative Humidity Range: 30% - 85 % (noncondensing)**  
**Ambient Temperature and humidity checked Using Traceable Calibration**  
**Control Co ([www.control3.com](http://www.control3.com))**  
**Model Number: ISO17025, S/N: 72544734, Certificate No: 4096-1713981**  
**Fisher Catalog No: 11-661-13**

**Validation Methodology:**

1. Run processor on regular overnight processing schedule using five practice blocks.
2. Run processor on weekend brain processing schedule using a practice block
3. Record oven temperature daily.

**Overnight processing schedule with test tissue: Date completed**\_\_\_\_\_  
**Ambient temp**\_\_\_\_\_, **Humidity**\_\_\_\_\_, **Oven Temp**\_\_\_\_\_

**Tissue processed**

**Approved  
By/date**


**Equipment Validation**  
**VIP5 Unit “C” (cont.)**

**Weekend brain processing schedule with test block:**  
**Date completed**\_\_\_\_\_, **Ambient Temp**\_\_\_\_\_, **Humidity**\_\_\_\_\_  
**Oven temp**\_\_\_\_\_

<b>Tissue Processed</b>	<b>Approved By/date</b>

**Equipment Validation  
ISOTEMP REFRIGERATOR**

**Model Number: Isotemp 20FREEFSA  
Serial Number: 01258874901150730**

**Installed: August 31, 2015**

**Clinical Engineering Number: MZB-REF-90  
Electrical Check (115Hz 60amps : 3phase)**

**Operating Conditions:**

**Ambient temperature range: -10 C to 10 C  
Temperature checked using traceable calibration control  
Model Number: FB50267  
S/N: 150103038  
Expiration date: 3/13/2017  
Fisher Catalog No: 06-664-11**

**Validation methodology:**

**Record temperatures for 6 consecutive working days using Traceable thermometer Serial #150103038 and compare with digital readout  $\pm$  4 C for consistency.**

<b>Date/tech</b>	<b>Temperature</b>	<b>Digital Readout</b>	<b>Ambient Temperature</b>	<b>Humidity</b>

**Supervisor**\_\_\_\_\_ **Date**\_\_\_\_\_  
**Notes:**

**Medical Director** \_\_\_\_\_ **Date**\_\_\_\_\_

**Equipment Validation  
Microtome**

**Manufacturer:** Reichert-Jung

**Model Number:** 2030

**Serial Number:** \_\_\_\_\_

**Clinical Engineering Number:** N/A – not electrical

**Installed:** June 27, 2008 (relocated)

**Specifications:** None stated

**Validation Methodology:** Cut slides from a current case for approval by a pathologist.

<b>Date/ Tech</b>	<b>Case#/ Tissue type</b>	<b>Approved by/date</b>

**Assigned to ( tech)**\_\_\_\_\_ **Date**\_\_\_\_\_

## UCSF Medical Center Whole Mount Blocks

Date: \_\_\_\_\_

Is the Peloris tissue processor comparable to the current tissue processing method as the Core lab?

(Circle one) YES NO

Does the tissue placed in the Peloris processor, run on the UCSF Bigs program, produce diagnostic slides?

(Circle one) YES NO

Approved by: \_\_\_\_\_

If NO, explain inadequacy:

Date	Test Case	Tissue Type	Core lab Results  (circle one)	Peloris Results  (circle one)	Comments	Suitable for UCSF MZ Histo lab?
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		

Histology Medical Director approval: \_\_\_\_\_ Date: \_\_\_\_\_

Supervisor review: \_\_\_\_\_ Date: \_\_\_\_\_



## UCSF Medical Center    Artisan Stain Validation

Date: \_\_\_\_\_

Special Stain: \_\_\_\_\_

Is this Artisan stain comparable to the current manual method and suitable for general use?  
(Circle one)    YES    NO

Approved by: \_\_\_\_\_

If NO, explain inadequacy:

Date	Test Case	Tissue Type	Manual Result (circle one)	Artisan Automation Result (Circle one)	Comments	Suitable to use Artisan for staining?
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		
			Inadequate Satisfactory	Inadequate Satisfactory		

## Immunohistochemistry Instrument Validation Testing

**Date:** 9-26-2011

**Instrument:** Leica Bond III **Serial Number** \_\_\_\_\_

**Circle one:**    New instrument                      Relocated instrument                      **Comparison Instrument** Leica Bond Max

**Installation Date**\_\_\_\_\_

**Instructions:** Run parallel tests on the existing instrument and the new, or relocated instrument

Approved by: \_\_\_\_\_


If NO, explain inadequacy:

[illegible]

## Automated Coverslipper **Demo Unit**

# Installation Checklist-Product # 6500

<b>Serial Number:</b>	Institution: UCSF	<b>Customer:</b>
Sakura Representative: Andy Anand	City/State: San Francisco, CA	
Installation Date:	<b>Customer Signature:</b>	

Check after completion= 	
<input type="checkbox"/>	Check area for space and proper ventilation.
<input type="checkbox"/>	Remove packing materials from coverslipper according to unpacking protocol Double check for materials inside unit. Remove tie wraps and protective foam from Coverslipper. Inspect for physical damage.
<input type="checkbox"/>	Place unit on counter and check for proper level adjust as needed.
<input type="checkbox"/>	Install activated carbon filters and/or duct adaptor.
<input type="checkbox"/>	Connect Power cord and plug into an 115V, 60Hz power receptacle.
<input type="checkbox"/>	Turn on power and allow initialization.
<input type="checkbox"/>	Prime instrument with clearing agent according to operators manual.
<input type="checkbox"/>	Position mounting medium on stage.
<input type="checkbox"/>	Load coverslips into unit
<input type="checkbox"/>	Install Link system if Prisma Stainer is to be connected.
<input type="checkbox"/>	Run test slides.
<input type="checkbox"/>	Train user on the following:
<input type="checkbox"/>	Review operating manual.
<input type="checkbox"/>	Using compatible clearing agents and mounting mediums.
<input type="checkbox"/>	Adjusting mounting medium dispensing.
<input type="checkbox"/>	Adjusting coverslip angle and speed.
<input type="checkbox"/>	Utilization of programming and program interface.
<input type="checkbox"/>	Loading slides and coverglass.
<input type="checkbox"/>	Recommended slide types.
<input type="checkbox"/>	Daily operation and maintenance.
<input type="checkbox"/>	Fill out warranty card.
<input type="checkbox"/>	Review service contract options and provide contact information.
<input type="checkbox"/>	Instruct customer on proper Technical support procedures.